

METHOD: Turbidimetric Immunoassay

Ordering information

Product	Package	Code No.
CRP-HSII LT R1 (Buffer)	R1: 2 x 40 mL	462-39493
CRP-HSII LT R2 (Antibody)	R2: 2 x 40 mL	468-39593
CRP-HS LT Calibrator Set HO	CAL: 5 conc. x 2 mL	460-25801

Measuring Temperature 37°C

Hitachi 902 / Hitachi 7070 (7060) / 911 (904) / 912

REAGENT	CRP-HSII LT
ANALYZE	
TEST	CRP
ASSAY CODE/ TIME	2 POINT END - 10
ASSAY POINTS	18 - 30 - 21 - 22
WAVELENGTH (SECONDARY / PRIMARY)	800 / 600
DILUTION	00400 / 0
SAMPLE VOLUME (NORMAL)	5 - 0 - 0
(DECREASE)	
(INCREASE)	
ABS. LIMIT	
PROZONE LIMIT	50 LOWER
REAGENT VOLUME R1	140 - 0 - () - 0
R2	0 - 0 - () - 0
R3	140 - 0 - () - 0
R4	0 - 0 - () - 0
CALIBRATION	
CALIBRATION TYPE	Spline (6)
NUMBER OF POINTS	6
SPAN POINT	5
CALIB. WEIGHT	0
AUTO CALIBRATION	
SD LIMIT	999,9
DUPLICATE LIMIT	500
SENSITIVITY LIMIT	0
S1 EXT LIMIT	-32000 / 32000
RANGE	
DATA MODE	
UNIT	mg/dL
INSTRUMENT FACTOR (Y = aX + b)	a = 1.0 b = 0.0
COMPENSATED LIMIT	
STANDARDS	CONC./POS./S.VOL./PREDIL./DIL.
1	0.00 / 1 / 5 / 0 / 0
2	*1 / *2 / 5 / 0 / 0
3	*1 / *2 / 5 / 0 / 0
4	*1 / *2 / 5 / 0 / 0
5	*1 / *2 / 5 / 0 / 0
6	*1 / *2 / 5 / 0 / 0
CALIBRATION LIST	
K-FACTOR	
SYSTEM PARAMETER	
LINEARITY CHECK (4 - 8)	
(9 -)	

*1: Input the assigned value of the calibrator (mg/L o. mg/dL).

*2: Input the position of the calibrator.

0111D1/vk