

Ordering information

Reagents	Package	Order No.
HDL-C L-Type R1	R1: 4 x 60 mL	412-72395
HDL-C L-Type R2	R2: 4 x 20 mL	412-72495
HDL-C L-Type R1	R1: 4 x 270 mL	418-72395
HDL-C L-Type R2	R2: 4 x 90 mL	418-72495
HDL-C Calibrator	CAL: 4 x for 3 mL	416-51095

AU400 / AU600 / AU640 (BECKMAN COULTER)

Reagent	HDL-C L-Type
Specific Test Parameters	
Test Name	*1 / HDL-C
Type	Serum
Sample: Volume / Dilution	2.5 µL / 0 µL
Reagents: R1 Volume / Dilution	210 µL / 0 µL
R2 Volume / Dilution	70 µL / 0 µL
Wavelength (Pri. / Sec.)	600 / 700
Method	END
Reaction slope	+
Measuring Point 1 (First / Last)	0 / 27
Measuring Point 2 (First / Last)	0 / 10
Linearity	[] %
No-Lag-Time	
Min OD / Max OD (L / H)	
Reagent OD Limit: First (L / H)	-2.0000 / 2.5000
Last (L / H)	-2.0000 / 2.5000
Dynamic Range (L / H)	0.02 - 4.65
Correlation Factor (A / B)	1.000000 / 0.000000
Calibration Specific	
Calibration Type	AB
Formula	1 (Y = aX + b)
Cuvette / Counts	2 / *1
Process	CONC
Calibrator	
Point 1	CAL# - OD - CONC *1 - [] - *2
Point 2	
Point 3	
Point 4	
Point 5	
Point 6	
Point 7	
1-Point Cal. Point	
MB Type Factor	

*1: To be inserted by the user

*2: Input the concentration of the calibrator