

## Ordering information

Reagents	Package	Order No.
HDL-C L-Type R1	R1: 4 x 60 mL	412-72395
HDL-C L-Type R2	R2: 4 x 20 mL	412-72495
HDL-C L-Type R1	R1: 4 x 270 mL	418-72395
HDL-C L-Type R2	R2: 4 x 90 mL	418-72495
HDL-C Calibrator	CAL: 4 x for 3 mL	416-51095

## Hitachi 902/ 911/ 912/ 904

REAGENT	HDL-C L-Type
<b>ANALYZE</b>	
TEST	HDL-C
ASSAY CODE/ TIME	2POINT END - 10
ASSAY POINTS	15 - 31 - 0 - 0
WAVELENGTH (SUB / MAIN)	700 / 600
DILUTION	/ 0
SAMPLE VOL. (NORM.)	3 - 0 - 0
(DEC.)	3 - 0 - 0
(INC.)	3 - 0 - 0
ABS. LIMIT	32000 / 32000 / INCREASE
PROZONE LIMIT	-32000 / 32000 / LOWER
REAGENT R1	270 - 0 - ( ) - 0
R2	0 - 0 - ( ) - 0
R3	90 - 0 - ( ) - 0
R4	0 - 0 - ( ) - 0
<b>CALIBRATION</b>	
CALIBRATION TYPE	LINEAR
NUMBER OF POINTS	2
SPAIN POINT	2
AUTO CALIBRATION	
SD LIMIT	999.9
DUPLICATE LIMIT	500
SENSITIVITY LIMIT	-99999 - 999999
S1 ABS LIMIT	-32000 / 32000
COMPENSATED LIMIT	
<b>RANGE</b>	
UNIT	mg/dL
INSTRUMENT FACTOR	(Y=aX+b)
a	1.0
b	0.0
<b>STANDARDS</b>	<b>Cal.Code/Conc./Pos./S.Vol./S.Dil./Dil.</b>
1	Water / 0.0 / - / 3 / 0 / 0
2	( * ) / *1 / *2 / 3 / 0 / 0
3	
4	
5	
6	
<b>CALIBRATION LIST</b>	
K-FACTOR	
<b>SYSTEM PARAMETER</b>	

\* : User-defined

\*1: Assigned value of the calibrator

\*2: Position of the calibrator